

Amend Section 80001 to read:

80001 DEFINITIONS

80001

(a) through (a)(9) (Continued)

(10) "Automated External Defibrillator" (AED), for adult community care means a light-weight, portable device used to administer an electric shock through the chest wall to the heart. Built-in computers assess the patient's heart rhythm, determine whether defibrillation (electrical shock) is needed and then administer the shock. Audible and/or visual prompts guide the user through the process. (Continued)

Authority Cited: Sections 1502.2, 1522.41(j), 1524(e), 1530, and 1530.9, Health and Safety Code.

Reference: Sections 1501, 1502, 1502(a)(7) and (8), 1502.2, 1502.5, 1503, 1503.5, 1505, 1507, 1508, 1509, 1511, 1520, 1522, 1524, ~~1524(e)~~, 1525, 1525.5, 1526, 1527, 1530, 1530.5, 1531, 1531.1, 1533, 1534, 1536.1, 1537, 1538.5, 1550, 1551, 1556, 1569.699(a), 1797.196, and 11834.11, Health and Safety Code; Sections 5453, 5458, 11006.9, and 17736(a) and (b), Welfare and Institutions Code; and 29 CFR 1910.1030.

Amend Section 80061 to read:

80061 REPORTS (Continued)

80061

(b) Upon the occurrence, during the operation of the facility, of any of the events specified in (1) below, a report shall be made to the licensing agency within the agency's next working day during its normal business hours. In addition, a written report containing the information specified in (2) below shall be submitted to the licensing agency within seven days following the occurrence of such event.

(1) Events reported shall include the following: (Continued)

(C) In an adult CCF, the use of an Automated External Defibrillator.

(D) (Continued)

(E) (Continued)

(F) (Continued)

(G) (Continued)

(H) (Continued)

(I) (Continued)

(J) (Continued)

Authority Cited: Sections 1530 and 1538.5, Health and Safety Code.

Reference: Sections 1501, 1507, 1531, ~~and~~ 1538.5, and 1797.196, Health and Safety Code.

Adopt Section 80075.1 to read:

80075.1 AUTOMATED EXTERNAL DEFIBRILLATORS (AEDs)

80075.1

- (a) In an adult community care facility, a licensee is permitted to maintain and operate an AED at the facility if all of the following requirements are met:
- (1) The licensee shall notify the licensing agency in writing that an AED is in the facility and will be used in accordance with all applicable federal and other state requirements.
 - (2) The AED shall be used in accordance with all applicable federal and other state requirements.
 - (3) The licensee shall maintain at the facility the following:
 - (A) A copy of the required physician's prescription for the AED.
 - (B) A training manual from an American Heart Association- or American Red Cross-recognized AED training class.
 - (C) A log of checks of operation of the AED containing the dates checked and the name of person checking.
 - (D) A copy of a valid AED operator's certificate for any employee(s) authorized by the licensee to operate the AED. If the certificate does not indicate that the AED training course completed complies with the standards of the American Heart Association or the American Red Cross, then other evidence so indicating must be available at the facility.
 - (E) A log of quarterly proficiency demonstrations for each holder of an AED operator's certificate who is authorized by the licensee to operate the AED. The log shall contain the dates of the demonstrations and the manner of demonstration.
 - (4) A supply kit shall be maintained at the facility and be readily available for use with the AED. The kit shall contain at least the following:
 - (A) A back-up battery set.
 - (B) An extra set of pads.
 - (C) A safety razor for shaving chest hair when necessary to apply the pads.
 - (D) A cardiovascular pulmonary resuscitation barrier (a face shield or mask) for protection from transmission of infectious disease.

- (E) Two pairs of unused medical examination gloves (latex or non-latex).
- (5) Use of an AED shall be reported as specified in Section 80061.

Authority Cited: Section 1530, Health and Safety Code.

Reference: Sections 1501, 1502, 1507, 1530, 1531, and 1797.196, Health and Safety Code.

Amend Section 82001 to read:

82001 DEFINITIONS

82001

For purposes of this chapter, the following definitions shall apply:

(a) (1) through (11) (Continued)

(12) "Automated External Defibrillator" (AED) means a light-weight, portable device used to administer an electric shock through the chest wall to the heart. Built-in computers assess the patient's heart rhythm, determine whether defibrillation (electrical shock) is needed and then administer the shock. Audible and/or visual prompts guide the user through the process. (Continued)

Authority Cited: Sections 1502.2, 1524(e), 1530, and 1530.9, Health and Safety Code.

Reference: Sections 1501, 1502, 1502(a)(7) and (8), 1502.2, 1503, 1503.5, 1505, 1507, 1508, 1509, 1511, 1520, 1522, 1524, ~~1524(e)~~, 1525, 1525.5, 1526, 1527, 1530, 1530.5, 1531, 1531.1, 1533, 1534, 1536.1, 1537, 1538.5, 1550, 1551, 1556, 1569.699(a), 1797.196, and 11834.11, Health and Safety Code; Sections 5453, 5458, 11006.9, and 17736(a) and (b), Welfare and Institutions Code; and 29 CFR 1910.1030.

Amend Section 82061 to read:

82061 REPORTS (Continued)

82061

(b) Upon the occurrence, during the hours the facility/center is providing services to the client, of any of the events specified in Section 82061(b)(1), a report shall be made to the licensing agency within the agency's next working day during its normal business hours. In addition, a written report containing the information specified in Section 82061(b)(2) shall be submitted to the licensing agency within seven days following the occurrence of such event.

(1) Events reported shall include the following: (Continued)

(C) The use of an Automated External Defibrillator.

~~(C)~~ (D) (Continued)

~~(D)~~ (E) (Continued)

~~(E)~~ (F) (Continued)

~~(F)~~ (G) (Continued)

~~(G)~~ (H) (Continued)

~~(H)~~ (I) (Continued)

(d) The items specified in Sections 82061(b)(1)(A) through (H) shall also be reported to the client's authorized representative, if any. (Continued)

Authority Cited: Sections 1530 and 1538.5, Health and Safety Code.

Reference: Sections 1501, 1507, 1520, 1531, ~~and~~ 1538.5, and 1797.196, Health and Safety Code.

Adopt Section 82075.2 to read:

82075.2 AUTOMATED EXTERNAL DEFIBRILLATORS (AEDs)

82075.2

- (a) A licensee is permitted to maintain and operate an AED at the facility if all of the following conditions are met:
- (1) The licensee shall notify the district office in writing that an AED is in the facility and will be used in accordance with all applicable federal and other state requirements.
 - (2) The AED shall be used in accordance with all applicable federal and other state requirements.
 - (3) The licensee shall maintain at the facility the following:
 - (A) A copy of the required physician's prescription for the AED.
 - (B) A training manual from an American Heart Association- or American Red Cross-recognized AED training class.
 - (C) A log of checks of operation of the AED containing the dates checked and the name of person checking.
 - (D) A copy of a valid AED operator's certificate for any employee(s) authorized by the licensee to operate the AED. If the certificate does not indicate that the AED training course completed complies with the standards of the American Heart Association or the American Red Cross, then other evidence so indicating must be available at the facility.
 - (E) A log of quarterly proficiency demonstrations for each holder of an AED operator's certificate who is authorized by the licensee to operate the AED. The log shall contain the dates of the demonstrations and the manner of demonstration.
 - (4) A supply kit shall be maintained at the facility and be readily available for use with the AED. The kit shall contain at least the following:
 - (A) A back-up battery set.
 - (B) An extra set of pads.
 - (C) A safety razor for shaving chest hair when necessary to apply the pads.
 - (D) A cardiovascular pulmonary resuscitation barrier (a face shield or mask) for protection from transmission of infectious disease.

- (E) Two pairs of unused medical examination gloves (latex or non-latex).
- (5) Use of an AED shall be reported as specified in Section 82061.

Authority Cited: Section 1530, Health and Safety Code.

Reference: Sections 1501, 1502, 1507, 1530, 1531, and 1797.196, Health and Safety Code.

Amend Section 87101 to read:

87101 DEFINITIONS

87101

(a) (1) through (6) (Continued)

(7) Automated External Defibrillator. "Automated External Defibrillator" (AED) means a light-weight, portable device used to administer an electric shock through the chest wall to the heart. Built-in computers assess the patient's heart rhythm, determine whether defibrillation (electrical shock) is needed and then administer the shock. Audible and/or visual prompts guide the user through the process. (Continued)

Authority Cited: Sections 1569.23(d), 1569.30, 1569.616(j), and 1569.698(c), Health and Safety Code.

Reference: 42 CFR 418.3; Sections 1569.1, 1569.2, 1569.5, 1569.10, 1569.145, 1569.15, 1569.153, 1569.157, 1569.158, 1569.17, 1569.19, 1569.191(e), 1569.193(a) and (c), 1569.20, 1569.21, 1569.23, 1569.30, 1569.31, 1569.312, 1569.38, 1569.44, 1569.47, 1569.54, 1569.616, 1569.699(a), 1569.73, 1569.74, 1569.82, 1797.196, 5350, and 7185 et seq., Health and Safety Code; and Sections 1800, 4615, and 4753, Probate Code.

Amend Section 87561 to read:

87561 REPORTS

87561

- (a) Each licensee shall furnish to the licensing agency such reports as the Department may require, including, but not limited to, the following:
- (1) A written report shall be submitted to the licensing agency and to the person responsible for the resident within seven days of the occurrence of any of the events specified in (A) through ~~(E)~~ below. This report shall include the resident's name, age, sex and date of admission; date and nature of event; attending physician's name, findings, and treatment, if any; and disposition of the case. (Continued)
- (C) The use of an Automated External Defibrillator.
- ~~(E)~~ (Continued)

Authority Cited: Sections 1569.30 and 1569.616(j), Health and Safety Code.

Reference: Sections 1569.1, 1569.2, 1569.30, 1569.31, 1569.312, ~~and~~ 1569.616, and 1797.196, Health and Safety Code.

Adopt Section 87575.2 to read:

87575.2 AUTOMATED EXTERNAL DEFIBRILLATORS (AEDs)

87575.2

- (a) A licensee is permitted to maintain and operate an AED at the facility if all of the following requirements are met:
- (1) The licensee shall notify the licensing agency in writing that an AED is in the facility and will be used in accordance with all applicable federal and other state requirements.
 - (2) The AED shall be used in accordance with all applicable federal and other state requirements.
 - (3) The licensee shall maintain at the facility the following:
 - (A) A copy of the required physician's prescription for the AED.
 - (B) A training manual from an American Heart Association- or American Red Cross-recognized AED training class.
 - (C) A log of checks of operation of the AED containing the dates checked and the name of person checking.
 - (D) A copy of a valid AED operator's certificate for any employee(s) authorized by the licensee to operate the AED. If the certificate does not indicate that the AED training course completed complies with the standards of the American Heart Association or the American Red Cross, then other evidence so indicating must be available at the facility.
 - (E) A log of quarterly proficiency demonstrations for each holder of an AED operator's certificate who is authorized by the licensee to operate the AED. The log shall contain the dates of the demonstrations and the manner of demonstration.
 - (4) A supply kit shall be maintained at the facility and be readily available for use with the AED. The kit shall contain at least the following:
 - (A) A back-up battery set.
 - (B) An extra set of pads.
 - (C) A safety razor for shaving chest hair when necessary to apply the pads.
 - (D) A cardiovascular pulmonary resuscitation barrier (a face shield or mask) for protection from transmission of infectious disease.

- (E) Two pairs of unused medical examination gloves (latex or non-latex).
- (5) Use of an AED shall be reported as specified in Section 87561.
- (6) Requests to Forego Resuscitative Measures, Advance Directives and Do-Not-Resuscitate Orders shall be observed as specified in Section 87575.1.

Authority Cited: Section 1569.30, Health and Safety Code.

Reference: Sections 1250, 1569.1, 1569.2, 1569.30, 1569.312, 1569.73, and 1797.196 Health and Safety Code.

Amend Section 87801 to read:

87801 DEFINITIONS

87801

- (a) The following definitions shall apply wherever the terms are used throughout Division 6, Chapter 8.5, except where specifically noted otherwise. (Continued)

(8) "Automated External Defibrillator" (AED) means a light-weight, portable device used to administer an electric shock through the chest wall to the heart. Built-in computers assess the patient's heart rhythm, determine whether defibrillation (electrical shock) is needed and then administer the shock. Audible and/or visual prompts guide the user through the process. (Continued)

Authority Cited: Section 1568.072, Health and Safety Code.

Reference: Sections 1568.01, 1568.02, 1568.03, 1568.061, 1568.064, ~~and~~ 1568.072, and 1797.196, Health and Safety Code.

Amend Section 87861 to read:

87861 REPORTS (Continued)

87861

(b) Upon the occurrence, during the operation of the facility, of any of the events specified in (1) below, a report shall be made to the Department on the same day or within the Department's next working day during its normal business hours. In addition, a written report containing the information specified in (2) below shall be submitted to the Department within seven days following the occurrence of such event.

(1) Events reported shall include the following: (Continued)

(C) The use of an Automated External Defibrillator.

(D) (Continued)

(E) (Continued)

(F) (Continued)

(G) (Continued)

(H) (Continued)

(I) (Continued)

(d) The items specified in (b)(1)(A) through (I) above shall also be reported to the resident's authorized representative, if any. (Continued)

Authority Cited: Section 1568.072, Health and Safety Code.

Reference: Sections 1568.072 and 1797.196, Health and Safety Code.

Adopt Section 87925 to read:

87925 AUTOMATED EXTERNAL DEFIBRILLATORS (AEDs)

87925

- (a) A licensee is permitted to maintain and operate an AED at the facility if all of the following requirements are met:
- (1) The licensee shall notify the licensing agency in writing that an AED is in the facility and will be used in accordance with all applicable federal and other state requirements.
 - (2) The AED shall be used in accordance with all applicable federal and other state requirements.
 - (3) The licensee shall maintain at the facility the following:
 - (A) A copy of the required physician's prescription for the AED.
 - (B) A training manual from an American Heart Association- or American Red Cross-recognized AED training class.
 - (C) A log of checks of operation of the AED containing the dates checked and the name of person checking.
 - (D) A copy of a valid AED operator's certificate for any employee(s) authorized by the licensee to operate the AED. If the certificate does not indicate that the AED training course completed complies with the standards of the American Heart Association or the American Red Cross, then other evidence so indicating must be available at the facility.
 - (E) A log of quarterly proficiency demonstrations for each holder of an AED operator's certificate who is authorized by the licensee to operate the AED. The log shall contain the dates of the demonstrations and the manner of demonstration.
 - (4) A supply kit shall be maintained at the facility and be readily available for use with the AED. The kit shall contain at least the following:
 - (A) A back-up battery set.
 - (B) An extra set of pads.
 - (C) A safety razor for shaving chest hair when necessary to apply the pads.
 - (D) A cardiovascular pulmonary resuscitation barrier (a face shield or mask) for protection from transmission of infectious disease.

- (E) Two pairs of unused medical examination gloves (latex or non-latex).
- (5) Any use of an AED shall be reported as specified in Section 87861(b).
- (6) Requests to Forego Resuscitative Measures, Advance Directives and Do-Not-Resuscitate Orders shall be observed as specified in Section 87924.

Authority Cited: Section 1568.072, Health and Safety Code.

Reference: Sections 1568.02, 1568.072, and 1797.196, Health and Safety Code.